

ABSTRACT OF THE DISCLOSURE

[0034] A deep brain stimulation lead system has a medical lead, or similar elongate medical insertion device, securable with a lead lock through a cannula slit, thereby allowing a lead to remain electrically operative and preventing movement of the lead during removal of a stylet, recording microelectrode, or cannula.